MEXPO INTERNATIONAL INC.
2828 Faber Street
Union City, CA 94587-1204

Declaration of Conformity
Rubber Dental Dam

PRODUCT DESCRIPTION

1. Product Name: Rubber Dental Dam
2. Product Classification: Class I under Medical Device Regulations (EU) 2017/745 Annex VIII Rule 1

ADDRESS:

Mexpo International Inc.
2828 Faber Street
Union City, CA 94587-1204
US & Canada Toll Free :
+1 (800) 838-8299
Tel : +1 (510) 489-6800
Fax : +1 (510) 489-3111
E-mail : blossomglo@aol.com
E-mail: tim@mexpo-glove.com

EC Representative:
Blossom Europe, S. L
Paseo de Recoletos, 37-41
28004 Madrid, Spain

We, **Mexpo International, Inc.** as the legal manufacturer declare under our sole responsibility that the medical devices listed below conform to the requirement of the medical device regulations (EU) 2017/745.

Rubber Dental Dam – BLOSSOM Color Green Mint

Size	UDI
5x5	10723860013320
6x6	10723860013344

Rubber Dental Dam – BLOSSOM Color Blue

Size	UDI
5x5	10723860014556
6x6	10723860013665

Rubber Dental Dam – BLOSSOM Color Light Blue

Size	UDI
5x5	10723860016550
6x6	10723860016666

Rubber Dental Dam – BLOSSOM Color Black

Size	UDI
5x5	10723860015553
6x6	10723860015669

Basic UDI-DI:72386011155PE

It is a non-invasive Class I device in accordance with Annex VIII of the Medical Device Regulation EU/2017/745, under Rule 1 as Rules 2, 3, and 4 do not apply, and meets all the provisions of the Medical Device Regulation 2017/745 which applies to them.

This declaration is supported by the Quality System approval ISO 13485 issued by TÜV SÜD America.
Conformity assessment procedure: In accordance with Medical Device Regulation 2017/745 Annex IX

The following person is responsible for the signature of this document:
Name and Address: Mexpo International, 2828 Faber Street, Union City, California 94587-1204, USA

Authorized Signature: _____ **Date:** July 12, 2021

Name of responsible Person: Tim Thai **Position:** President

Mexpo International Inc.