META BIOMED CO., LTD.
Osong factory): 270, Osongsaengmyeong1-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, Korea
T) +82-43-216-0433 F) +82-43-217-1988
Email: info@meta-biomed.com Home Page: www.meta-biomed.com

EC Declaration of Conformity

according to the Medical Devices Directive 93/42/EEC as amended by Directive 2007/47/EC

For the following medical device :

Product Name : Root Canal Sealer

Brand Name : ADSEAL (detailed model names attached)

Classification : Class Ⅱa by Rule 8 of Annex Ⅸ, Council Directive 93/42/EEC as amended by Directive 2007/47/EC

Manufacturer's Name : META BIOMED CO., LTD.

Manufacturer's Address : 270, Osongsaengmyeong1-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, Korea

is herewith the conformity with the requirements set out in the Council Directive 93/42/EEC of 14 June 1993 as amended by Directive 2007/47/EC of the European Parliament and of the Council of 5 September 2007 concerning Medical Devices declared, For the evaluation of the conformity with this Directive, the following standards have been applied:

MDD 93/42/EEC, EN ISO 13485:2016, MEDDEV 2.12-1 Rev.8, MEDDEV 2.7.1 Rev.4, MEDDEV 2.12.2 Rev.2, EN ISO 14971:2019, ISO/TR 24971:2020, EN 1041:2008, EN ISO 15223-1:2016, EN 1641:2009, EN 62366-1:2015, IEC TR 62366-2:2016, ISO/TR 20416:2020, EN ISO 6876:2012, ADA/ANSI 57:2000, EN ISO 7405:2018, EN ISO 10993-1:2009, EN ISO 10993-3:2014, EN ISO 10993-5:2009, EN ISO 10993-6:2009, EN ISO 10993-10:2013, EN ISO 10993-11:2018, ASTM F1980-21, EN ISO 14644-1:2015, EN ISO 14644-2:2015, EN ISO 14644-3:2019, EN ISO 14644-4:2001, EN ISO 14644-5:2004, KS I ISO 14644-6:2017, EN ISO 14644-7:2004, EN ISO 14644-8:2013, ISO 14698-1:2003, ISO 14698-2:2003, ISTA 1A 2016

is subject to the procedure set out in Annex Ⅱ(excluding sect-4) of Directive 93/42/EEC under the supervision of Notified Body Number 1639, SGS Belgium NV, Noorderlaan 87, BE-2030 Antwerpen, Belgium.

GMDN Code : 36095(Endodontic filling/sealing material)

EC Certificate Number : KR19/81826259

Responsible for making this declaration is the :

■ Manufacturer: META BIOMED CO., LTD.

European Representative is:

■ Meta Biomed Europe GmbH, Wiesenstr. 35, 45473 Mülheim an der Ruhr, Germany

■ Tel:+49-208-30991910, Fax: +49-208-30991999

We, the manufacturer's declaration, declare and ensure with sole responsibility, that the Medical Device meet the provisions of Council Directive 93/42/EEC (Medical Device Directive) which apply to them.

Person responsible for making this declaration:

Name, Surname : Suk Song Oh

Position/Title : PRESIDENT

Place: Cheongju-si **Date:** 05 December, 2022

Suk Song Oh

(Company stamp and legal signature)

META BIOMED CO., LTD.
Osong factory): 270, Osongsaengmyeong1-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, Korea
T) +82-43-216-0433 F) +82-43-217-1988
Email: info@meta-biomed.com Home Page: www.meta-biomed.com

(Attachment: Model name list)

No	Brand Name	Model Name
1	ADSEAL	ADSEAL
2		ADSEAL 4.5